



**University of
Zurich**^{UZH}

**Zurich Open Repository and
Archive**

University of Zurich
University Library
Strickhofstrasse 39
CH-8057 Zurich
www.zora.uzh.ch

Year: 2015

Peri-implant infections of oral biofilm etiology

Belibasakis, Georgios N ; Charalampakis, Georgios ; Bostanci, Nagihan ; Stadlinger, Bernd

Abstract: Biofilms are complex microbial communities that grow on various surfaces in nature. The oral microbiota tend to form polymicrobial biofilms, particularly on the hard mineralized surfaces of teeth, which may impact on oral health and disease. They can cause inflammation of the adjacent tooth-supporting (periodontal) tissues, leading to destructive periodontal disease and tooth loss. The emergence of osseointegrated dental implants as a restorative treatment option for replacing missing teeth has also brought along new artificial surfaces within the oral cavity, on which oral bacteria can form biofilms. As in the case of natural teeth, biofilms on implant surfaces may also trigger infection and cause inflammatory destruction of the peri-implant tissue (i.e. peri-implantitis). While there are strong similarities in the composition of the mixed microbial flora between periodontal and peri-implant infections, there are also a few distinctive differences. The immunological events underlying the pathogenesis of peri-implant infections are qualitatively similar, yet more extensive, compared to periodontal infections, resulting in a faster progression of tissue destruction. This chapter summarizes the current knowledge on the microbiology and immunology of peri-implant infections, including findings from the peri-implant crevicular fluid, the inflammatory exudate of the peri-implant tissue. Moreover, it discusses the diagnosis and current approaches for the treatment of oral infections.

DOI: https://doi.org/10.1007/978-3-319-11038-7_4

Posted at the Zurich Open Repository and Archive, University of Zurich

ZORA URL: <https://doi.org/10.5167/uzh-102760>

Book Section

Accepted Version

Originally published at:

Belibasakis, Georgios N; Charalampakis, Georgios; Bostanci, Nagihan; Stadlinger, Bernd (2015). Peri-implant infections of oral biofilm etiology. In: Donelli, Gianfranco. Biofilm-based Healthcare-associated Infections - Volume I. Cham: Springer, 69-84.

DOI: https://doi.org/10.1007/978-3-319-11038-7_4

Chapter:

Peri-implant infections of oral biofilm etiology

**Georgios N. Belibasakis ^{1*}, Georgios Charalampakis ², Nagihan Bostanci ³,
Bernd Stadlinger ⁴**

1. Section of Oral Microbiology and Immunology, Institute of Oral Biology, Center of Dental Medicine, University of Zürich, Plattenstrasse 11, 8032 Zürich, Switzerland

2. Department of Oral Microbiology and Immunology, Institute of Odontology, Gothenburg University, The Sahlgrenska Academy at University of Gothenburg, Box 450 Gothenburg 405 30, Sweden.

3. Section of Oral Translational Research, Institute of Oral Biology, Center of Dental Medicine, University of Zürich, Plattenstrasse 11, 8032 Zürich, Switzerland

4. Clinic of Oral Surgery, Center of Dental Medicine, University of Zürich, Plattenstrasse 11, 8032 Zürich, Switzerland

* Corresponding author. E-mail: george.belibasakis@zzm.uzh.ch

Summary: Biofilms are complex microbial communities that grow on various surfaces in nature. The oral microbiota tend to form polymicrobial biofilms, particularly on the hard mineralized surfaces of teeth, which may impact on oral health and disease. They can cause inflammation of the adjacent tooth-supporting (periodontal) tissues, leading to destructive periodontal disease and tooth loss. The emergence of osseointegrated dental implants as a restorative treatment option for replacing missing teeth has also brought along new artificial surfaces within the oral cavity, on which oral bacteria can form biofilms. As in the case of natural teeth, biofilms on implant surfaces may also trigger infection and cause inflammatory destruction of the peri-implant tissue (i.e. peri-implantitis). While there are strong similarities in the composition of the mixed microbial flora between periodontal and peri-implant infections, there are also a few distinctive differences. These are likely to expand as the application of metagenomics and metatranscriptomics becomes increasingly popular in the analysis of oral ecology. The immunological events underlying the pathogenesis of peri-implant infections are qualitatively similar, yet more extensive, compared to periodontal infections, resulting in a faster progression of tissue destruction. This chapter summarizes the current knowledge on the microbiology and immunology of peri-implant infections, including findings from the peri-implant crevicular fluid, the inflammatory exudate of the peri-implant tissue. Moreover, it discusses the diagnosis and current approaches for the treatment of oral infections.

1.1 Biofilms in the oral cavity

The oral cavity consists of both soft mucosal tissue surfaces and hard dental tissues all bathed in constantly secreted saliva. This anatomical niche of the human body constitutes a dynamic ecosystem that is continuously colonized by commensal microorganisms, which are collectively defined as the oral microbial flora. They have evolved along with the host, while their survival is tightly dependent on their capacity to use the available nutrients for their growth and their adaptability to the host's innate and adaptive immune system. It is estimated that the diversity of oral microbiota accounts for more than 700 different species, with at least 100 species populating the oral cavity of a given individual (Aas et al. 2005; Paster et al. 2001). The oral bacteria rarely grow in single planktonic form, but they naturally form biofilm communities with each other on the tooth surface. Biofilms exhibit a very high level of structural and functional bacterial organization, whereby the individual bacterial constituents communicate with each other by finely tuned molecular processes (also defined as “quorum sensing”) (Huang et al. 2014). Biofilms demonstrate much more virulent characteristics compared to bacteria in planktonic state, as they exhibit altered gene expression patterns and are less penetrable by neutrophils, antibodies, or antimicrobial factors (Schaudinn et al. 2009), even by a factor of 500 (Costerton et al. 1995). Clinically, the “dental plaque” forming on the tooth surface holds the full properties of a biofilm (Marsh 2003).

Changes in the local microenvironment may cause shifts in the composition of the biofilm microflora, giving leeway to certain bacterial species to overgrow, enhance their virulence properties and eventually become opportunistic pathogens. Such species may be found at low numbers in health and can become pathogenic only when the newly established conditions

permit them so. This is the principle drive for the “ecological plaque hypothesis”, the predominant theory that explains the aetiology of the polymicrobial oral diseases as a disturbance of the relationship between the resident oral microbiota and the response of the host that they populate (Marsh 2003; Marsh and Devine 2011). Dysbiotic biofilms can endure the host response and concomitantly exploit the inflammatory host response, in a manner that propagates the magnitude of the inflammatory tissue destruction, as is the case of bone loss in periodontitis (Hajishengallis and Lamont 2012; Hajishengallis 2014).

1.2 Biofilms and oral disease

Dental caries and periodontal diseases are the two main and highly prevalent oral diseases, both caused by biofilms growing on the tooth surface. Dental caries manifests essentially as the demineralization of the hard dental tissues (namely enamel and dentine), by acids generated due to the fermentation of dietary carbohydrates by the biofilms grown on the tooth surface. Its incidence has increased with sugar consumption and it is among the most prevalent infectious diseases in the industrialised world. If dental caries remains untreated, the biofilm-associated bacteria can eventually invade into the deeper soft dental pulp tissue, causing pulpitis, and subsequently tooth necrosis. Periodontal diseases are a major group of biofilm-associated oral diseases that destroy the tooth-supporting (periodontal) tissues as a result of excessive inflammatory response of the juxtaposed gingival tissue. Aetiologically, they are attributed to polymicrobial biofilms accumulated on the tooth surface, and particularly under the gingival margin (subgingival). The inflammation can be contained within the gingival tissue (gingivitis), manifests as swelling and bleeding of the gingiva, symptoms that are easily identifiable by the patient. This condition affects virtually the whole global population, and is reversible once the

biofilm is removed from the tooth surface and proper oral hygiene is instilled. Persistence and progression of an exacerbated the inflammatory response can destroy the deeper periodontal tissues, namely the periodontal ligament that links it to the tooth surface and the supporting alveolar bone. The disease has now progressed to periodontitis which, apart from the loss of supporting bone and periodontal ligament, is also characterised by the formation of deep periodontal pockets. If left untreated, periodontitis will result in exfoliation of the affected tooth, impairing the chewing function and compromising the aesthetic appearance. Periodontitis is the main cause of tooth loss in the industrialised world, and is perhaps the most prevalent inflammatory infectious disease in human adults, affecting approximately 1/3 of population. One of the restorative treatment options for replacing teeth missing due to periodontitis is dental implants.

2.1 Dental implants and comparison to natural teeth

Dental implants are titanium-based screw-like devices that are surgically installed into the jaw bone, in the place of one or more missing teeth. Thereafter, a transmucosal abutment is adapted onto the implant, mediating the connection to the final prosthetic restoration that is visible in the patient's oral cavity. Hence, the patient's functional and aesthetic needs are re-established. The titanium surface is biologically "accepted" by the surrounding bone tissue, and forms a connection known as "osseointegration". This titanium-bone relationship possesses the essential functional properties required to support the replacement of a missing tooth.

There is merit at this stage to define the main dissimilarities between periodontal and peri-implant tissues, in order to better understand peri-implant infections, or diseases (Figure 1A and 1B). In the case of dental implants, the main difference with natural teeth is the absence of

periodontal ligament, thus necessitating direct interface between the bone and the implant surface (Heitz-Mayfield and Lang 2010). Instead, the collagen fibers of the submucosal connective tissue are arranged parallel to the implant surface, thus forming a “collar”. Consequently, the resulting in the peri-implant crevice is deeper than the gingival crevice of natural teeth, resulting in a weaker physical barrier against bacterial invasion. Apart from very restricted mobility, the lack of the periodontal ligament also means restricted blood supply. Hence, the delivery of cells of the immune system, needed to tackle the early stages of bacterial infection, is rather compromised. Collectively, these characteristics may render dental implants more susceptible to endogenous infection, than natural teeth. An exposed dental implant surface is also prone to microbial colonization. Thus biofilms can as well form on implants, with potential detrimental effects on the health of the surrounding peri-implant tissues.

3.1 Peri-implant infections: classification and diagnosis

Failures of dental implant function are classified either as early, or as late ones (Listgarten 1997; Tabanella et al. 2009). The early ones refer to incomplete osseointegration following surgical installation, and may be attributed to early loading, surgical contamination, poor compatibility of the material, or inefficient healing due to systemic disease. In late failures, the normal function of an already osseointegrated implant is disrupted, resulting from chronic infection of the peri-implant tissues.

Peri-implant mucositis is characterized by biofilm-induced inflammation localized on the soft peri-implant mucosa, but with no evidence of destruction of the supporting bone (Figure 1C). Progression of the inflammation may lead to gradual destruction the bone, manifesting as peri-implantitis (Figure 1D). Peri-implant mucositis and peri-implantitis are analogous to

gingivitis and periodontitis of natural teeth (Heitz-Mayfield and Lang 2010). Mucositis occurs in approximately 80% of patients with dental implants, and 50% of the implants. The prevalence of peri-implantitis has varied reportedly from 28%-56% among patients, and 12%-43% among implants (Zitzmann and Berglundh 2008).

To-date the diagnosis of peri-implant diseases is based on clinical and radiographic criteria (Mombelli and Lang 1998; Kao et al. 1997). While mucositis is characterized by inflammation, erythema and bleeding of the tissue particularly during examination, peri-implantitis exhibits additionally a peri-implant pocket which is greater than 4 mm, potentially suppuration, and a characteristic “saucer-“ or “crater-shaped” bone destruction around the implant, which is revealed radiographically (Heitz-Mayfield 2008). Peri-implantitis is becoming a pathological entity of growing concern among clinicians because of its aggressive pattern, and in certain cases rapid reach of the terminal stage i.e. implant loss. It is a biological complication around dental implants, characterized by profound inflammation of the tissues surrounding an implant in function with progressive loss of supporting bone (Lindhe and Meyle 2008).

4.1 Peri-implantitis as a biofilm-initiated disease

Despite the early introduction of the term “peri-implantitis” (Levignac 1965; Lindhe and Meyle 2008), the infectious nature of the disease was documented almost 2 decades later (Rams and Link 1983; Rams et al. 1991). Treatment with osseointegrated implants was introduced in fully edentulous patients, and in such individuals there was no biological rationale to consider post-osseointegration infections, since periodontitis- associated bacteria were to be automatically ‘removed’ from the oral cavity together with the extracted teeth. Late implant failure was explained during many years by overloading or excess loading. However, a recent systematic

review (Naert et al. 2012) clarified that no association between overload and peri-implant bone loss could be found in the absence of peri-implant inflammation. Indeed, research should address critical questions with regard to aetiopathogenesis of peri-implantitis, which in turn would guide evidence-based treatment.

Given the non-shedding surface of the dental implant in the oral cavity, it is easy to understand why biofilm formation on the implant is an inadvertent process. The mouth provides not only a portal of entry for bacteria but also an inherent environment for bacterial colonization and growth. In a similar fashion to the tooth, by the time the dental implant is exposed in the oral cavity, it is covered by an acquired pellicle layer i.e. an organic stratum mainly consisting of proteins, glycoproteins and lipids. The pellicle triggers early bacterial colonization by providing receptors for the adhesins of specific species of oral bacteria. Early colonizers are *Streptococci* and *Actinomyces* species (Nakazato et al. 1989; Li et al. 2004) and create the preconditions for the accumulation of late-colonizing bacteria such as *Fusobacterium* and *Prevotella* species (Hannig 1997; Aas et al. 2005). The bacterial colonization of the surface starts already 30 min after implant insertion, and these early bacterial species can be found as part of the mixed biofilm community on the implant surface even several months later (Furst et al. 2007). The mature biofilm can eventually detach with dispersal and spread further, a critical stage for bacterial dissemination and consequent colonization of deeper tissue sites.

In terms of initial (i.e. 4 weeks) subgingival colonization, the frequency of detection of different species is similar between natural teeth and implants. Nevertheless, the colonization pattern on implants appears to be initially slower than on natural teeth (Quirynen et al. 2005), given the pristine surfaces of the implant and the lack of the desired indigenous microbiota. The bacterial composition of the biofilm formed on implants closely resembles that of the

neighboring teeth (Botero et al. 2005; Salvi et al. 2008). This leads us to postulate that the oral microbial flora, and especially that of neighboring natural teeth, acts as a “reservoir” for the biofilms that build-up around implants.

4.2 Peri-implant microflora resembling periodontal microflora

The peri-implant microflora in health consists mainly of Gram-positive cocci and non-motile bacilli, and a limited number Gram-negative anaerobic species, resembling gingival health (Mombelli et al. 1987; Bower et al. 1989). The switch to peri-implant mucositis is associated with increased presence of cocci, motile bacilli and spirochetes, comparable to gingivitis (Pontoriero et al. 1994). The transition to peri-implantitis is characterized with the emergence of Gram-negative, motile, and anaerobic species that are commonly found in periodontitis (Mombelli et al. 1987; Mombelli and Decaillet 2011). An interesting finding is that the microbial composition of peri-implant pockets in partially edentulous patients resembles that of the neighboring periodontal pockets, while the presence of *Porphyromonas gingivalis*, *Tannerella forsythia* and *Treponema denticola* can be higher in peri-implantitis (Botero et al. 2005; Hultin et al. 2002; Shibli et al. 2008). Hence, the qualitative composition of the microbial flora of peri-implantitis-associated biofilms is in concordance with periodontitis. A representative microscopy image of a submucosal biofilm sample obtained from a site with peri-implantitis is provided in Figure 2, whereby the diversity of morphotypes and taxa can be depicted. Finally, a finding of further interest is that some bacteria retrieved from peri-implantitis biofilms, most often *Prevotella intermedia/nigrescens* and *Streptococcus constellatus*, may display *in vitro* resistance to one or more standard antibiotic treatments (Rams et al. 2013).

4.3 Distinct peri-implant microflora

By use of molecular techniques in microbiological analyses we have nowadays appreciated the breadth of microbial diversity in the subgingival/submucosal biofilms. Though logical it may sound that implants and neighboring teeth share similar microbiota since they share a similar ecological niche i.e. interdental space, emerging evidence suggests that they could be microbiologically distinct from each other (Kumar et al. 2012; Dabdoub et al. 2013; Heuer et al. 2012). By use of broad-range PCR techniques (Heuer et al. 2012) and pyrosequencing (Kumar et al. 2012; Dabdoub et al. 2013) it was demonstrated that the peri-implant microbiome was distinct from the periodontal microbiome. Given the distinct topography and immunological characteristics of the implant compared to the tooth, the two ecosystems could be regarded divergent. This could explain why teeth and implants may harbour diverse bacterial lineages.

A number of microorganisms have been identified in peri-implantitis that are less regularly detected in periodontitis. These include, but not restricted to, *Staphylococcus aureus*, *Staphylococcus epidermidis*, *Enterobacter aerogenes*, *Enterobacter cloace*, *Escherichia coli*, *Helicobacter pylori*, *Pseudomonas* spp, as well as *Candida* spp fungi (Renvert et al. 2007; Rams and Link 1983; Persson and Renvert 2013; Belibasakis 2014; Rams et al. 1991; Leonhardt et al. 2003; Leonhardt et al. 2002; Leonhardt et al. 1999). *S. aureus* is a versatile human pathogen discussed extensively in orthopedics as the leading etiologic agent of implant infection and of the associated osteomyelitis (Arciola 2009). It has a wide array of virulence factors, including up to 21 different adhesins or Microbial Surface Components Recognizing Adhesive Matrix Molecules (MSCRAMMs) and cytotoxins (Patti et al. 1994; Darouiche et al. 1997; Speziale et al. 2009). For this reasons, it has attracted attention over recent years as a specific pathogen for peri-implantitis, distinct to periodontitis. The presence of *S. aureus* shortly after dental implant

insertion can be confirmed even one year later (Salvi et al. 2008), while a recent microbiological study revealed by checkerboard methodology that significantly higher counts of *S. aureus* and *Staphylococcus anaerobius* were detected at implants with peri-implantitis than healthy implants (Persson and Renvert 2013).

Aerobic Gram negative bacilli (AGNB) include two wide and distinct categories: i) bacteria that ferment lactose and belong to the large family *Enterobacteriaceae* (i.e. *E. coli*, *Enterobacter*, *Klebsiella*, *Citrobacter*), and ii) non-enteric rods that do not ferment lactose (i.e. *Pseudomonas aeruginosa*, *Acinetobacter baumannii*). In a retrospective investigation of peri-implantitis cases (Charalampakis et al. 2012) culture analysis demonstrated the presence of aerobic Gram-negative bacilli at moderately heavy or heavy growth in 18.6% of patients with peri-implantitis. However, the microbial burden could not fully correspond to peri-implant disease severity. AGNB have been detected in previous studies both in peri-implantitis cases (Botero et al. 2005; Alcoforado et al. 1991; Leonhardt et al. 1999; Rosenberg et al. 1991), as well as in healthy implants (Nowzari et al. 2008b; Leonhardt et al. 1999; Nowzari et al. 2008a). Thus, the true role of AGNB in the etiology of peri-implant infections remains unclear.

4.4 Bacteria in the implant-abutment interface

A bacterial leakage along the implant-abutment interface has been also discussed in the literature (do Nascimento et al. 2008; do Nascimento et al. 2009; Gross et al. 1999; Callan et al. 2005). Given the fact that the interface includes microgaps ranging from 10 to 100 μm , we cannot exclude the scenario that microorganisms of 2 μm or less in diameter penetrate the passive fit between the implant components. Poor fit of attached components, inadequate torque, geometry of the implant platforms between the various implant systems, poor stability and micro-

movements in the deeper inner portions of the system may enhance the extent of bacterial leakage (Binon 2000; Aloise et al. 2010). Such microgaps may function as ‘nests’ for anaerobic or microaerophilic bacteria to be protected from host defense mechanisms and persist for extended periods. However, the hermetic closure of a contaminated small compartment would serve as entombment, creating unfavorable conditions for bacteria to grow. Thus, the risk for peri-implant infection around two-part implant systems because of microbial leakage per se should be considered minimal.

4.5 Effect of implant surface on biofilm

Rough implant surfaces were introduced in the dental market in order to enhance the rate of osseointegration. However, the implant surface roughness has a significant impact on the quantity and the quality of the plaque formed. A rough surface structure characterized by grooves and pits may provide the bacteria with ‘protected’ areas, inaccessible to conventional mechanical removal (Renvert et al. 2008). Other surface characteristics that enhance initial bacterial adhesion are high wettability and great surface free energy (Teughels et al. 2006). A recent *in vivo* study on the biofilm structure formed on titanium discs with different surface characteristics revealed different microbial patterns (Charalampakis et al. 2014). By SEM analysis it was demonstrated that the discs representing the moderately rough surfaces (Osseospeed™, TiOblast™, Experimental surface) harbored a complex biofilm with tight intercellular bacterial bindings, whereas the discs with the turned surface hosted a biofilm that presented a pattern of spread bacteria forming less clusters. The study concluded that variations in the biofilm pattern may be associated with the different surface characteristics of titanium discs.

However, there is limited and contradictory evidence on the impact of implant surface on peri-implantitis. Some studies have found a positive correlation between smooth surface and peri-implant health (Astrand et al. 2004; Esposito et al. 2007), whereas others failed to find a correlation between type of implant surface and marginal bone loss (Gotfredsen and Karlsson 2001; Wennstrom et al. 2004). Nevertheless, it is also shown that surface characteristics of the abutments may not influence biofilm formation, or the extent and cellular composition of the inflammatory lesion (Zitzmann et al. 2002). Accordingly, no implant system or surface type was found to be superior in terms of marginal bone preservation (Abrahamsson and Berglundh 2009).

Last but not least, like in natural teeth, implant surfaces are immediately populated by salivary mucoproteins, which are required for the adhesion of bacteria (Kolenbrander et al. 2010). These are genetically defined in each individual, and may coat the surfaces of both natural teeth and implants, before being recognized by the same bacterial species. It is therefore tempting to postulate that potential differences in bacterial adhesion due to surface microstructure may partially be equilibrated by the mediating salivary pellicle (Busscher et al. 2010). Hence, given the inevitable mediation of the pellicle in bacterial adhesion, implant surface characteristics may not notably affect the initial stages of biofilm formation.

5.1 Histopathological events in peri-implant infections

Like in the case of natural teeth, the development of a biofilm on the implant surface is an igniting factor of the inflammatory response of the surrounding peri-implant tissues. Peri-implant mucositis is characterized by inflammation that is confined to epithelium, connective tissue loss, microvascular changes (Sanz et al. 1990), and increased infiltration of leukocytes (Zitzmann et al. 2002; Zitzmann et al. 2004). The sequence of inflammatory events are that take place in peri-

implant mucositis are similar to those in gingivitis, but potentially of a larger extent than gingivitis (Ericsson et al. 1992; Berglundh et al. 1991; Berglundh et al. 1992). The switch to peri-implantitis is accompanied by a further influx of inflammatory cells than in the affected area of the peri-implant mucosa, that now expands to reach the bone tissue (Gualini and Berglundh 2003; Talarico et al. 1997; Lindhe et al. 1992), while a large number of osteoclasts form onto the bone surface and initiate bone resorption (Carcuac et al. 2013).

5.2 Immune responses to biofilm in peri-implant infections

The histopathological events associated with peri-implant infections have been characterized over the past two decades. Nevertheless, there are still pending questions regarding the molecular regulatory events underlying these described processes. Peri-implant mucosal tissue biopsies, as well as inflammatory tissue exudates are suitable biological material to investigate in depth the molecular events associated with peri-implant diseases. In that respect, diseased tissue obtained from peri-implantitis sites is shown to exhibit higher expression of several mediators of inflammation, including pro-inflammatory cytokines interleukin (IL)-6, IL-8 and tumor necrosis factor (TNF)- α , compared to healthy or peri-implant mucositis sites (Venza et al. 2010; Kontinen et al. 2006; Duarte et al. 2009b). Although tissue biopsies can give an actual view of the undergoing molecular events with the tissue, the invasiveness of the collection process makes it almost impossible to use this material on a regular basis. In contrast, the collection process of the inflammatory exudate of the peri-implant tissues, namely the peri-implant crevicular fluid (PICF), is much simpler and non-invasive. This topic is discussed further in the next section.

5.3 Peri-implant crevicular fluid as a reservoir of inflammatory mediators

The PICF is the inflammatory exudate of the peri-implant sulcus or crevice, which is the tight anatomical deepening formed between the implant surface and the peri-implant mucosa. This niche can be converted into a peri-implant pocket as peri-implantitis progresses, fostering a submucosal biofilm. Similarly to gingival crevicular fluid (GCF) which bathes natural teeth, PICF is the outcome of increased permeability of the vessels within the underlying connective tissue, as an inflammatory response to the growing biofilm. Although, the molecular characterization of PICF is at its infancy compared to GCF, it is already known to contain several serum and locally produced molecules, such as tissue breakdown products, inflammatory mediators and antibodies directed against the bacteria of the biofilm (Adonogianaki et al. 1995). Therefore, analysis of the PICF might be suitable to evaluate the inflammatory status of peri-implant tissues, in a quantitative manner (Kaklamanos and Tsalikis 2002; Belibasakis 2014).

In healthy implant tissues, the diffusion of PICF is rather passive and slow. However, its volume amount is significantly increased at a given site once biofilm-induced inflammation is established (i.e. peri-implant mucositis). Human experimentally induced peri-implant mucositis studies have elegantly showed that both PICF volume and protein content increases by the end of the 3-week period of plaque accumulation protocol (Salvi et al. 2012). Despite the strong clinical similarities among human experimental gingivitis and peri-implant mucositis, the latter is presented with a more pronounced inflammatory response to biofilm accumulation (Salvi et al. 2012; Pontoriero et al. 1994; Schierano et al. 2008). Since the composition of PICF is modified along with the histopathological changes during the course of progressive inflammation, its molecular analysis may support the early detection of clinically undetectable changes.

Pro-inflammatory cytokines have been the primary candidates to be investigated in PICF, due to their central role in triggering the inflammatory process, and the good amount on knowledge already derives in studies on GCF. It was confirmed that higher concentrations of TNF- α , IL-17 and IL-1 β are present in PICF collected from peri-implantitis-affected sites, compared to healthy controls (Ataoglu et al. 2002; Curtis et al. 1997; Darabi et al. 2013; Vieira et al. 2013; Severino et al. 2011), whereas the opposite was the case for anti-inflammatory cytokine IL-10 (Casado et al. 2013). More recently multiplex cytokine arrays have been applied in the analysis of PICF, allowing for a broader simultaneous screening of multiple inflammatory cytokines and chemokines. In one such study, 12 markers were assessed in both peri-implant health and disease, including granulocyte macrophage colony-stimulating factor (GM-CSF), IL-1 β , IL-2, IL-4, IL-5, IL-6, IL-7, IL-8, IL-10, IL-12, interferon (IFN)- γ and TNF- α (Fonseca et al. 2014). In line with previous findings, there were no differences with regards to IL-6, IL-8, IL-10 (Severino et al. 2011), whereas the levels of IL-1 β were significantly higher at peri-implantitis, compared to mucositis sites. Further studies also showed that peri-implantitis treatment reduced the PICF levels of IL-1 β (Basseti et al. 2013) and TNF- α (de Mendonca et al. 2009; Duarte et al. 2009a). Studies have also demonstrated that single nucleotide polymorphisms (SNP) in the IL-1 gene may hold an increased risk for the development of peri-implantitis, particularly when combined with smoking (Andreiotelli et al. 2008; Bormann et al. 2010). However, the presence of a specific SNP does not necessarily translate into higher levels of IL-1 β in PICF (Lachmann et al. 2007; Jansson et al. 2005; Melo et al. 2012). Although there is good evidence of the involvement of IL-1 β as a crucial mediator of the host inflammatory response in peri-implant tissues (Murata et al. 2002; Salvi et al. 2010; Ataoglu et al. 2002), this needs to be complemented with further interventional studies, whereby IL-1 β inhibitors are part of the treatment. Moreover,

other members of the IL-1 family, such as IL-18, which may display differential regulation from IL-1 β , should also be considered in PICF (Hamed et al. 2009; Bostanci et al. 2009).

Matrix metalloproteinases (MMPs) are proteolytic enzymes that with a strongly documented role in collagen degradation in various disease processes, including periodontitis (Sorsa et al. 2006). Their involvement in peri-implant tissue destruction has also received attention, and a number of studies have demonstrated that MMP levels in PICF from peri-implantitis sites are elevated compared to healthy sites, and that their enzymatic activity increases with disease severity (Kivela-Rajamaki et al. 2003a; Kivela-Rajamaki et al. 2003b; Ozcakir-Tomruk et al. 2012; Xu et al. 2008; Teronen et al. 1997), while successful treatment results in reduction (Wohlfahrt et al. 2014; Salvi et al. 2012). Moreover, studies to determine whether MMP-8 is useful prognostic marker of peri-implantitis, have shown that sites higher high PICF levels of MMP-8 are at greater risk for progressive bone loss (Arakawa et al. 2012).

As bone resorption is the hallmark of peri-implantitis, the regulation of osteoclastogenesis and osteogenesis-associated markers have also been studied in PICF. Similar to findings in GCF obtained from sites with periodontitis (Belibasakis and Bostanci 2012), there is increasing evidence showing the association of the receptor activator of nuclear factor kappaB ligand (RANKL) - osteoprotegerin (OPG) system with the occurrence and severity of peri-implantitis (Arikan et al. 2008; Arikan et al. 2011; Sarlati et al. 2010; Guncu et al. 2012; Duarte et al. 2009a; Duarte et al. 2009b; Rakic et al. 2013). A few studies looking into other markers of bone metabolism, such as osteocalcin, reported that these are higher in PICF from peri-implant mucositis compared to healthy sites, whereas no differences were observed between peri-implantitis and either mucositis or health. Hence, elevated levels of osteocalcin in PICF may

reflect increased local bone turnover around implants, rather than severe bone resorption (Murata et al. 2002).

6.1 Treatment of peri-implant infections

The development of peri-implantitis shows comparable features to the development of periodontitis (Heitz-Mayfield and Lang 2010). Clinical treatment of peri-implantitis is performed by various means and there is currently no consensus on an official standard of care. Therapy generally aims at the settlement of inflammatory peri-implant processes and the preservation of hard and soft tissues, as evaluated by reduced bleeding on probing and reduced probing depth or stable radiographic bone level, respectively (Heitz-Mayfield and Mombelli 2014). Four phases of treatment are suggested in order to enable the successful treatment of peri-implantitis: 1) pre-treatment phase (oral hygiene, prosthodontic aspects), 2) surgical access (mucoperiosteal flap, bone substitute with or without membranes), 3) post-operative anti-infective control (systemic antibiotics, chlorhexidine rinses), 4) maintenance care (3-6 months maintenance) (Heitz-Mayfield and Mombelli 2014).

Supportive periodontal therapy is seen as means to reduce the likelihood of an onset of peri-implantitis (Salvi and Zitzmann 2014). Pre-surgical therapy of peri-implantitis should include measures of oral hygiene. These mostly result in a reduction of mucositis, by targeting the disruption of the associated biofilm (Mishler and Shiau 2014). Also air abrasive powders and laser treatment have been applied for the reduction of biofilms on implant surfaces (Mishler and Shiau 2014; Schwarz et al. 2013). Further, the mechanical debridement of the implant surface should be performed, using instruments that cause little trauma to the surface, antiseptics and possibly antibiotics (Lang et al. 1997). Surgical therapy enables the debridement of granulation

tissue within a peri-implant defect and the possible performance of an implantoplasty by diamond burs, smoothening implant threads and structured implant surfaces to a polished state. Different surface topographies show differences in their susceptibility for peri-implant inflammation. During surgical therapy, the mechanical implantoplasty modifies the surface state, depending on the implant material (e.g. titanium grade 4) and the prior surface treatment (e.g. sandblasted, acid etched). This causes differences in post-surgical peri-implant bone formation (Albouy et al. 2011).

An additional regenerative step may be the application of bone replacement substances with or without membranes, aiming at peri-implant hard tissue regeneration. The desired aim is the re-osseointegration of a previously biofilm-covered surface. This regenerative therapy may be influenced by the peri-implant defect morphology and the absence or presence of keratinized mucosa (Schwarz et al. 2010). The type of prosthodontic restoration also influences the surgery. The operative site can be easily assessed through the removal of screw-retained implant crowns. In case of cemented crowns, debridement and implantoplasty may be limited to the buccal and approximal areas. A meta-analysis of treatment outcomes identified four surgical procedures, namely a) access flap and debridement, b) surgical resection, c) regeneration with bone grafts, and d) guided bone regeneration. A reduction in probing depth of 2-3 mm and a mean 2 mm radiographic bone gain is described for regenerative procedures. Most of these analyses have follow-ups of 1-2 years (Chan et al. 2013). Outcomes for regenerative approaches are described to vary the most.

Following surgical therapy, the onset of oral hygiene measures by the patient is important for long-term treatment success. It is important to maintain periodontal care over many years, as peri-implantitis may occur 5-10 years after implant placement and may reoccur at any time after

treatment. Bone loss >2 mm post-prosthetic treatment and bleeding on probing could be as an indicator of peri-implantitis (Klinge 2012). A probing depth threshold of 5 mm without any bleeding on probing may predict the cessation of bone loss and the successful outcome of peri-implantitis therapy (Klinge 2012; Heitz-Mayfield and Mombelli 2014) In case of an onset of bone loss, implant mobility will follow. In such cases implant removal, the debridement of the peri-implant defect, tissue regeneration and possibly re-implantation will follow. A clinical example of a peri-implantitis case is provided in Figure 3.

7.1 Concluding remarks on peri-implant infections

It is evident that dental implant surfaces provide a suitable substrate for the growth of oral biofilms, in a similar manner as natural teeth. This is not without consequences, as the uncontrolled biofilm formation due to inefficient oral hygiene will eventually cause inflammation of the surrounding tissues in the form of peri-implant mucositis, and potentially lead to tissue destruction, manifesting as peri-implantitis. It is clear that the microbiota of the peri-implant biofilms derives from the various micro-ecological niches in the oral cavity, including neighboring teeth, periodontal pockets, mucosal tissues and saliva. Although in principle the mixed microflora of peri-implant infections resembles that of periodontal ones, a number of non-typical oral taxa are more frequently found in peri-implantitis than periodontitis. Such are staphylococci, AGNB (e.g. enterobacteria and *Pseudomonas* spp) and *Candida* spp. The application of metagenomics in the analyses of biofilm samples is also likely to reveal specific microbial signatures in peri-implant infections. With regards to the pathogenesis of peri-implant infections the qualitative composition and sequence of the underlying immunological events resemble those of periodontal infections, but their magnitude is greater resting in a more

aggressive progression of the disease. Hence peri-implant infections are “contemporary” infections caused by oral biofilms etiology. They have emerged along with the continuous application of dental implants in restorative dentistry. While they are currently being treated in a similar philosophy as periodontal diseases, there is a need for reconsideration of their distinctive differences. This could lead to highly specialized therapeutic protocols, optimized for peri-implantitis.

References

- Aas JA, Paster BJ, Stokes LN, Olsen I, Dewhirst FE (2005) Defining the normal bacterial flora of the oral cavity. *J Clin Microbiol* 43 (11):5721-5732
- Abrahamsson I, Berglundh T (2009) Effects of different implant surfaces and designs on marginal bone-level alterations: a review. *Clin Oral Implants Res* 20 Suppl 4:207-215
- Adonogianaki E, Mooney J, Wennstrom JL, Lekholm U, Kinane DF (1995) Acute-phase proteins and immunoglobulin G against *Porphyromonas gingivalis* in peri-implant crevicular fluid: a comparison with gingival crevicular fluid. *Clin Oral Implants Res* 6 (1):14-23
- Albouy JP, Abrahamsson I, Persson LG, Berglundh T (2011) Implant surface characteristics influence the outcome of treatment of peri-implantitis: an experimental study in dogs. *J Clin Periodontol* 38 (1):58-64. doi:10.1111/j.1600-051X.2010.01631.x
- Alcoforado GA, Rams TE, Feik D, Slots J (1991) Microbial aspects of failing osseointegrated dental implants in humans. *J Parodontol* 10 (1):11-18
- Aloise JP, Curcio R, Laporta MZ, Rossi L, da Silva AM, Rapoport A (2010) Microbial leakage through the implant-abutment interface of Morse taper implants in vitro. *Clin Oral Implants Res* 21 (3):328-335

- Andreiotelli M, Koutayas SO, Madianos PN, Strub JR (2008) Relationship between interleukin-1 genotype and peri-implantitis: a literature review. *Quintessence Int* 39 (4):289-298
- Arakawa H, Uehara J, Hara ES, Sonoyama W, Kimura A, Kanyama M, Matsuka Y, Kuboki T (2012) Matrix metalloproteinase-8 is the major potential collagenase in active peri-implantitis. *J Prosthodont Res* 56 (4):249-255
- Arciola CR (2009) New concepts and new weapons in implant infections. *Int J Artif Organs* 32 (9):533-536
- Arikan F, Buduneli N, Kutukculer N (2008) Osteoprotegerin levels in peri-implant crevicular fluid. *Clin Oral Implants Res* 19 (3):283-288
- Arikan F, Buduneli N, Lappin DF (2011) C-telopeptide pyridinoline crosslinks of type I collagen, soluble RANKL, and osteoprotegerin levels in crevicular fluid of dental implants with peri-implantitis: a case-control study. *Int J Oral Maxillofac Implants* 26 (2):282-289
- Astrand P, Engquist B, Dahlgren S, Grondahl K, Engquist E, Feldmann H (2004) Astra Tech and Branemark system implants: a 5-year prospective study of marginal bone reactions. *Clin Oral Implants Res* 15 (4):413-420
- Ataoglu H, Alptekin NO, Haliloglu S, Gursel M, Ataoglu T, Serpek B, Durmus E (2002) Interleukin-1beta, tumor necrosis factor-alpha levels and neutrophil elastase activity in peri-implant crevicular fluid. *Clin Oral Implants Res* 13 (5):470-476
- Bassetti M, Schar D, Wicki B, Eick S, Ramseier CA, Arweiler NB, Sculean A, Salvi GE (2013) Anti-infective therapy of peri-implantitis with adjunctive local drug delivery or photodynamic therapy: 12-month outcomes of a randomized controlled clinical trial. *Clin Oral Implants Res*. doi:10.1111/clr.12155
- Belibasakis GN (2014) Microbiological and immuno-pathological aspects of peri-implant diseases. *Arch Oral Biol* 59 (1):66-72

- Belibasakis GN, Bostanci N (2012) The RANKL-OPG system in clinical periodontology. *J Clin Periodontol* 39 (3):239-248
- Berglundh T, Lindhe J, Ericsson I, Marinello CP, Liljenberg B, Thomsen P (1991) The soft tissue barrier at implants and teeth. *Clin Oral Implants Res* 2 (2):81-90
- Berglundh T, Lindhe J, Marinello C, Ericsson I, Liljenberg B (1992) Soft tissue reaction to de novo plaque formation on implants and teeth. An experimental study in the dog. *Clin Oral Implants Res* 3 (1):1-8
- Binon PP (2000) Implants and components: entering the new millennium. *Int J Oral Maxillofac Implants* 15 (1):76-94
- Bormann KH, Stuhmer C, Z'Graggen M, Kokemoller H, Rucker M, Gellrich NC (2010) IL-1 polymorphism and periimplantitis. A literature review. *Schweizer Monatsschrift fur Zahnmedizin* = *Revue mensuelle suisse d'odonto-stomatologie* = *Rivista mensile svizzera di odontologia e stomatologia* / *SSO* 120 (6):510-520
- Bostanci N, Emingil G, Saygan B, Turkoglu O, Atilla G, Curtis MA, Belibasakis GN (2009) Expression and regulation of the NALP3 inflammasome complex in periodontal diseases. *Clin Exp Immunol* 157 (3):415-422
- Botero JE, Gonzalez AM, Mercado RA, Olave G, Contreras A (2005) Subgingival microbiota in peri-implant mucosa lesions and adjacent teeth in partially edentulous patients. *J Periodontol* 76 (9):1490-1495
- Bower RC, Radny NR, Wall CD, Henry PJ (1989) Clinical and microscopic findings in edentulous patients 3 years after incorporation of osseointegrated implant-supported bridgework. *J Clin Periodontol* 16 (9):580-587
- Busscher HJ, Rinastiti M, Siswomihardjo W, van der Mei HC (2010) Biofilm formation on dental restorative and implant materials. *J Dent Res* 89 (7):657-665

- Callan DP, Cobb CM, Williams KB (2005) DNA probe identification of bacteria colonizing internal surfaces of the implant-abutment interface: a preliminary study. *J Periodontol* 76 (1):115-120
- Carcuac O, Abrahamsson I, Albouy JP, Linder E, Larsson L, Berglundh T (2013) Experimental periodontitis and peri-implantitis in dogs. *Clin Oral Implants Res* 24 (4):363-371
- Casado PL, Canullo L, de Almeida Filardy A, Granjeiro JM, Barboza EP, Leite Duarte ME (2013) Interleukins 1beta and 10 expressions in the periimplant crevicular fluid from patients with untreated periimplant disease. *Implant dentistry* 22 (2):143-150
- Chan HL, Lin GH, Suarez F, Maceachern M, Wang HL (2013) Surgical Management of Peri-implantitis: A Systematic Review and Meta-analysis of Treatment Outcomes. *J Periodontol*
- Charalampakis G, Leonhardt A, Rabe P, Dahlen G (2012) Clinical and microbiological characteristics of peri-implantitis cases: a retrospective multicentre study. *Clin Oral Implants Res* 23 (9):1045-1054
- Charalampakis G, Ramberg P, Dahlen G, Berglundh T, Abrahamsson I (2014) Effect of cleansing of biofilm formed on titanium discs. *Clin Oral Implants Res*
- Costerton JW, Lewandowski Z, Caldwell DE, Korber DR, Lappin-Scott HM (1995) Microbial biofilms. *Annu Rev Microbiol* 49:711-745
- Curtis DA, Kao R, Plesh O, Finzen F, Franz L (1997) Crevicular fluid analysis around two failing dental implants: a clinical report. *J Prosthodont* 6 (3):210-214
- Dabdoub SM, Tsigarida AA, Kumar PS (2013) Patient-specific analysis of periodontal and peri-implant microbiomes. *J Dent Res* 92 (12 Suppl):168S-175S
- Darabi E, Kadkhoda Z, Amirzargar A (2013) Comparison of the Levels of Tumor Necrosis Factor-alpha and Interleukin-17 in Gingival Crevicular Fluid of Patients with Peri-implantitis and a Control Group with Healthy Implants. *Iranian journal of allergy, asthma, and immunology* 12 (1):75-80. doi:012.01/ijaa.7580

- Darouiche RO, Landon GC, Patti JM, Nguyen LL, Fernau RC, McDevitt D, Greene C, Foster T, Klima M (1997) Role of Staphylococcus aureus surface adhesins in orthopaedic device infections: are results model-dependent? J Med Microbiol 46 (1):75-79
- de Mendonca AC, Santos VR, Cesar-Neto JB, Duarte PM (2009) Tumor necrosis factor-alpha levels after surgical anti-infective mechanical therapy for peri-implantitis: a 12-month follow-up. J Periodontol 80 (4):693-699
- do Nascimento C, Barbosa RE, Issa JP, Watanabe E, Ito IY, Albuquerque RF, Jr. (2008) Bacterial leakage along the implant-abutment interface of premachined or cast components. International journal of oral and maxillofacial surgery 37 (2):177-180
- do Nascimento C, Barbosa RE, Issa JP, Watanabe E, Ito IY, de Albuquerque Junior RF (2009) Use of checkerboard DNA-DNA hybridization to evaluate the internal contamination of dental implants and comparison of bacterial leakage with cast or pre-machined abutments. Clin Oral Implants Res 20 (6):571-577
- Duarte PM, de Mendonca AC, Maximo MB, Santos VR, Bastos MF, Nociti FH (2009a) Effect of anti-infective mechanical therapy on clinical parameters and cytokine levels in human peri-implant diseases. J Periodontol 80 (2):234-243
- Duarte PM, de Mendonca AC, Maximo MB, Santos VR, Bastos MF, Nociti Junior FH (2009b) Differential cytokine expressions affect the severity of peri-implant disease. Clin Oral Implants Res 20 (5):514-520
- Ericsson I, Berglundh T, Marinello C, Liljenberg B, Lindhe J (1992) Long-standing plaque and gingivitis at implants and teeth in the dog. Clin Oral Implants Res 3 (3):99-103
- Esposito M, Murray-Curtis L, Grusovin MG, Coulthard P, Worthington HV (2007) Interventions for replacing missing teeth: different types of dental implants. Cochrane Database Syst Rev (4):CD003815

- Fonseca FJ, Moraes Junior M, Lourenco EJ, Teles Dde M, Figueredo CM (2014) Cytokines expression in saliva and peri-implant crevicular fluid of patients with peri-implant disease. *Clin Oral Implants Res* 25 (2):e68-72
- Furst MM, Salvi GE, Lang NP, Persson GR (2007) Bacterial colonization immediately after installation on oral titanium implants. *Clin Oral Implants Res* 18 (4):501-508
- Gotfredsen K, Karlsson U (2001) A prospective 5-year study of fixed partial prostheses supported by implants with machined and TiO₂-blasted surface. *J Prosthodont* 10 (1):2-7
- Gross M, Abramovich I, Weiss EI (1999) Microleakage at the abutment-implant interface of osseointegrated implants: a comparative study. *Int J Oral Maxillofac Implants* 14 (1):94-100
- Gualini F, Berglundh T (2003) Immunohistochemical characteristics of inflammatory lesions at implants. *J Clin Periodontol* 30 (1):14-18
- Guncu GN, Akman AC, Gunday S, Yamalik N, Berker E (2012) Effect of inflammation on cytokine levels and bone remodelling markers in peri-implant sulcus fluid: a preliminary report. *Cytokine* 59 (2):313-316
- Hajishengallis G (2014) Immunomicrobial pathogenesis of periodontitis: keystones, pathobionts, and host response. *Trends Immunol* 35 (1):3-11
- Hajishengallis G, Lamont RJ (2012) Beyond the red complex and into more complexity: the polymicrobial synergy and dysbiosis (PSD) model of periodontal disease etiology. *Mol Oral Microbiol* 27 (6):409-419
- Hamed M, Belibasakis GN, Cruchley AT, Rangarajan M, Curtis MA, Bostanci N (2009) *Porphyromonas gingivalis* culture supernatants differentially regulate interleukin-1 β and interleukin-18 in human monocytic cells. *Cytokine* 45 (2):99-104
- Hannig M (1997) Transmission electron microscopic study of in vivo pellicle formation on dental restorative materials. *Eur J Oral Sci* 105 (5 Pt 1):422-433

- Heitz-Mayfield LJ (2008) Peri-implant diseases: diagnosis and risk indicators. *J Clin Periodontol* 35 (8 Suppl):292-304. doi:10.1111/j.1600-051X.2008.01275.x
- Heitz-Mayfield LJ, Lang NP (2010) Comparative biology of chronic and aggressive periodontitis vs. peri-implantitis. *Periodontol* 2000 53:167-181
- Heitz-Mayfield LJ, Mombelli A (2014) The therapy of peri-implantitis: a systematic review. *Int J Oral Maxillofac Implants* 29 Suppl:325-345
- Heuer W, Kettenring A, Stumpp SN, Eberhard J, Gellermann E, Winkel A, Stiesch M (2012) Metagenomic analysis of the peri-implant and periodontal microflora in patients with clinical signs of gingivitis or mucositis. *Clinical oral investigations* 16 (3):843-850
- Huang R, Li M, Gregory RL (2014) Bacterial interactions in dental biofilm. *Virulence* 2 (5):435-444
- Hultin M, Gustafsson A, Hallstrom H, Johansson LA, Ekfeldt A, Klinge B (2002) Microbiological findings and host response in patients with peri-implantitis. *Clin Oral Implants Res* 13 (4):349-358
- Jansson H, Hamberg K, De Bruyn H, Bratthall G (2005) Clinical consequences of IL-1 genotype on early implant failures in patients under periodontal maintenance. *Clin Implant Dent Relat Res* 7 (1):51-59
- Kaklamanos EG, Tsalikis L (2002) A review on peri-implant crevicular fluid assays potential in monitoring and predicting peri-implant tissue responses. *J Int Acad Periodontol* 4 (2):49-59
- Kao RT, Curtis DA, Murray PA (1997) Diagnosis and management of peri-implant disease. *J Calif Dent Assoc* 25 (12):872-880
- Kivela-Rajamaki M, Maisi P, Srinivas R, Tervahartiala T, Teronen O, Husa V, Salo T, Sorsa T (2003a) Levels and molecular forms of MMP-7 (matrilysin-1) and MMP-8 (collagenase-2) in diseased human peri-implant sulcular fluid. *J Periodontal Res* 38 (6):583-590

- Kivela-Rajamaki MJ, Teronen OP, Maisi P, Husa V, Tervahartiala TI, Pirila EM, Salo TA, Mellanen L, Sorsa TA (2003b) Laminin-5 gamma2-chain and collagenase-2 (MMP-8) in human peri-implant sulcular fluid. *Clin Oral Implants Res* 14 (2):158-165
- Klinge B (2012) Peri-implant marginal bone loss: an academic controversy or a clinical challenge? *European journal of oral implantology* 5 Suppl:S13-19
- Kolenbrander PE, Palmer RJ, Jr., Periasamy S, Jakubovics NS (2010) Oral multispecies biofilm development and the key role of cell-cell distance. *Nat Rev Microbiol* 8 (7):471-480
- Kontinen YT, Lappalainen R, Laine P, Kitti U, Santavirta S, Teronen O (2006) Immunohistochemical evaluation of inflammatory mediators in failing implants. *The International journal of periodontics & restorative dentistry* 26 (2):135-141
- Kumar PS, Mason MR, Brooker MR, O'Brien K (2012) Pyrosequencing reveals unique microbial signatures associated with healthy and failing dental implants. *J Clin Periodontol* 39 (5):425-433
- Lachmann S, Kimmerle-Muller E, Axmann D, Scheideler L, Weber H, Haas R (2007) Associations between peri-implant crevicular fluid volume, concentrations of crevicular inflammatory mediators, and composite IL-1A -889 and IL-1B +3954 genotype. A cross-sectional study on implant recall patients with and without clinical signs of peri-implantitis. *Clin Oral Implants Res* 18 (2):212-223. doi:10.1111/j.1600-0501.2006.01322.x
- Lang NP, Mombelli A, Tonetti MS, Bragger U, Hammerle CH (1997) Clinical trials on therapies for peri-implant infections. *Ann Periodontol* 2 (1):343-356
- Leonhardt A, Bergstrom C, Lekholm U (2003) Microbiologic diagnostics at titanium implants. *Clin Implant Dent Relat Res* 5 (4):226-232
- Leonhardt A, Grondahl K, Bergstrom C, Lekholm U (2002) Long-term follow-up of osseointegrated titanium implants using clinical, radiographic and microbiological parameters. *Clin Oral Implants Res* 13 (2):127-132

- Leonhardt A, Renvert S, Dahlen G (1999) Microbial findings at failing implants. *Clin Oral Implants Res* 10 (5):339-345
- Levignac J (1965) [Periimplantation osteolysis- periimplantosis - periimplantitis]. *Rev Fr Odontostomatol* 12 (8):1251-1260
- Li J, Helmerhorst EJ, Leone CW, Troxler RF, Yaskell T, Haffajee AD, Socransky SS, Oppenheim FG (2004) Identification of early microbial colonizers in human dental biofilm. *J Appl Microbiol* 97 (6):1311-1318
- Lindhe J, Berglundh T, Ericsson I, Liljenberg B, Marinello C (1992) Experimental breakdown of peri-implant and periodontal tissues. A study in the beagle dog. *Clin Oral Implants Res* 3 (1):9-16
- Lindhe J, Meyle J (2008) Peri-implant diseases: Consensus Report of the Sixth European Workshop on Periodontology. *J Clin Periodontol* 35 (8 Suppl):282-285
- Listgarten MA (1997) Clinical trials of endosseous implants: issues in analysis and interpretation. *Annals of periodontology / the American Academy of Periodontology* 2 (1):299-313.
doi:10.1902/annals.1997.2.1.299
- Marsh PD (2003) Are dental diseases examples of ecological catastrophes? *Microbiology* 149 (Pt 2):279-294
- Marsh PD, Devine DA (2011) How is the development of dental biofilms influenced by the host? *J Clin Periodontol* 38 Suppl 11:28-35
- Melo RF, Lopes BM, Shibli JA, Marcantonio E, Jr., Marcantonio RA, Galli GM (2012) Interleukin-1beta and interleukin-6 expression and gene polymorphisms in subjects with peri-implant disease. *Clin Implant Dent Relat Res* 14 (6):905-914
- Mishler O, Shiau HJ (2014) Management of Peri-Implant Disease: a current appraisal. *J Evid Based Dent Pract*
- Mombelli A, Decaillet F (2011) The characteristics of biofilms in peri-implant disease. *J Clin Periodontol* 38 Suppl 11:203-213

- Mombelli A, Lang NP (1998) The diagnosis and treatment of peri-implantitis. *Periodontol* 2000 17:63-76
- Mombelli A, van Oosten MA, Schurch E, Jr., Land NP (1987) The microbiota associated with successful or failing osseointegrated titanium implants. *Oral Microbiol Immunol* 2 (4):145-151
- Murata M, Tatsumi J, Kato Y, Suda S, Nunokawa Y, Kobayashi Y, Takeda H, Araki H, Shin K, Okuda K, Miyata T, Yoshie H (2002) Osteocalcin, deoxypyridinoline and interleukin-1 β in peri-implant crevicular fluid of patients with peri-implantitis. *Clin Oral Implants Res* 13 (6):637-643
- Naert I, Duyck J, Vandamme K (2012) Occlusal overload and bone/implant loss. *Clin Oral Implants Res* 23 Suppl 6:95-107
- Nakazato G, Tsuchiya H, Sato M, Yamauchi M (1989) In vivo plaque formation on implant materials. *Int J Oral Maxillofac Implants* 4 (4):321-326
- Nowzari H, Botero JE, DeGiacomo M, Villacres MC, Rich SK (2008a) Microbiology and cytokine levels around healthy dental implants and teeth. *Clin Implant Dent Relat Res* 10 (3):166-173
- Nowzari H, Yi K, Chee W, Rich SK (2008b) Immunology, microbiology, and virology following placement of NobelPerfect scalloped dental implants: analysis of a case series. *Clin Implant Dent Relat Res* 10 (3):157-165
- Ozcakir-Tomruk C, Chiquet M, Mericske-Stern R (2012) Tenascin-C and matrix metalloproteinase-9 levels in crevicular fluid of teeth and implants. *Clin Implant Dent Relat Res* 14 (5):672-681
- Paster BJ, Boches SK, Galvin JL, Ericson RE, Lau CN, Levanos VA, Sahasrabudhe A, Dewhirst FE (2001) Bacterial diversity in human subgingival plaque. *J Bacteriol* 183 (12):3770-3783
- Patti JM, Allen BL, McGavin MJ, Hook M (1994) MSCRAMM-mediated adherence of microorganisms to host tissues. *Annu Rev Microbiol* 48:585-617
- Persson GR, Renvert S (2013) Cluster of Bacteria Associated with Peri-Implantitis. *Clin Implant Dent Relat Res*
- Pontoriero R, Tonelli MP, Carnevale G, Mombelli A, Nyman SR, Lang NP (1994) Experimentally induced peri-implant mucositis. A clinical study in humans. *Clin Oral Implants Res* 5 (4):254-259

- Quirynen M, Vogels R, Pauwels M, Haffajee AD, Socransky SS, Uzel NG, van Steenberghe D (2005) Initial subgingival colonization of 'pristine' pockets. *J Dent Res* 84 (4):340-344
- Rakic M, Lekovic V, Nikolic-Jakoba N, Vojvodic D, Petkovic-Curcin A, Sanz M (2013) Bone loss biomarkers associated with peri-implantitis. A cross-sectional study. *Clin Oral Implants Res* 24 (10):1110-1116
- Rams TE, Degener JE, van Winkelhoff AJ (2013) Antibiotic resistance in human peri-implantitis microbiota. *Clin Oral Implants Res*. doi:10.1111/clr.12160
- Rams TE, Link CC, Jr. (1983) Microbiology of failing dental implants in humans: electron microscopic observations. *J Oral Implantol* 11 (1):93-100
- Rams TE, Roberts TW, Feik D, Molzan AK, Slots J (1991) Clinical and microbiological findings on newly inserted hydroxyapatite-coated and pure titanium human dental implants. *Clin Oral Implants Res* 2 (3):121-127
- Renvert S, Lindahl C, Renvert H, Persson GR (2008) Clinical and microbiological analysis of subjects treated with Branemark or AstraTech implants: a 7-year follow-up study. *Clin Oral Implants Res* 19 (4):342-347
- Renvert S, Roos-Jansaker AM, Lindahl C, Renvert H, Rutger Persson G (2007) Infection at titanium implants with or without a clinical diagnosis of inflammation. *Clin Oral Implants Res* 18 (4):509-516
- Rosenberg ES, Torosian JP, Slots J (1991) Microbial differences in 2 clinically distinct types of failures of osseointegrated implants. *Clin Oral Implants Res* 2 (3):135-144
- Salvi GE, Aglietta M, Eick S, Sculean A, Lang NP, Ramseier CA (2012) Reversibility of experimental peri-implant mucositis compared with experimental gingivitis in humans. *Clin Oral Implants Res* 23 (2):182-190

- Salvi GE, Franco LM, Braun TM, Lee A, Rutger Persson G, Lang NP, Giannobile WV (2010) Pro-inflammatory biomarkers during experimental gingivitis in patients with type 1 diabetes mellitus: a proof-of-concept study. *J Clin Periodontol* 37 (1):9-16
- Salvi GE, Furst MM, Lang NP, Persson GR (2008) One-year bacterial colonization patterns of *Staphylococcus aureus* and other bacteria at implants and adjacent teeth. *Clin Oral Implants Res* 19 (3):242-248
- Salvi GE, Zitzmann NU (2014) The effects of anti-infective preventive measures on the occurrence of biologic implant complications and implant loss: a systematic review. *Int J Oral Maxillofac Implants* 29 Suppl:292-307
- Sanz M, Newman MG, Nachnani S, Holt R, Stewart R, Flemmig T (1990) Characterization of the subgingival microbial flora around endosteal sapphire dental implants in partially edentulous patients. *Int J Oral Maxillofac Implants* 5 (3):247-253
- Sarlati F, Sattari M, Gazar AG, Rafsenjani AN (2010) Receptor activator of nuclear factor kappa B ligand (RANKL) levels in peri-implant crevicular fluid. *Iran J Immunol* 7 (4):226-233
- Schaudinn C, Gorur A, Keller D, Sedghizadeh PP, Costerton JW (2009) Periodontitis: an archetypical biofilm disease. *J Am Dent Assoc* 140 (8):978-986
- Schierano G, Pejrone G, Brusco P, Trombetta A, Martinasso G, Preti G, Canuto RA (2008) TNF-alpha TGF-beta2 and IL-1beta levels in gingival and peri-implant crevicular fluid before and after de novo plaque accumulation. *J Clin Periodontol* 35 (6):532-538
- Schwarz F, Hegewald A, John G, Sahm N, Becker J (2013) Four-year follow-up of combined surgical therapy of advanced peri-implantitis evaluating two methods of surface decontamination. *J Clin Periodontol* 40 (10):962-967
- Schwarz F, Sahm N, Schwarz K, Becker J (2010) Impact of defect configuration on the clinical outcome following surgical regenerative therapy of peri-implantitis. *J Clin Periodontol* 37 (5):449-455

- Severino VO, Napimoga MH, de Lima Pereira SA (2011) Expression of IL-6, IL-10, IL-17 and IL-8 in the peri-implant crevicular fluid of patients with peri-implantitis. *Arch Oral Biol* 56 (8):823-828. doi:10.1016/j.archoralbio.2011.01.006
- Shibli JA, Melo L, Ferrari DS, Figueiredo LC, Faveri M, Feres M (2008) Composition of supra- and subgingival biofilm of subjects with healthy and diseased implants. *Clin Oral Implants Res* 19 (10):975-982
- Sorsa T, Tjaderhane L, Konttinen YT, Lauhio A, Salo T, Lee HM, Golub LM, Brown DL, Mantyla P (2006) Matrix metalloproteinases: contribution to pathogenesis, diagnosis and treatment of periodontal inflammation. *Annals of medicine* 38 (5):306-321. doi:10.1080/07853890600800103
- Speziale P, Pietrocola G, Rindi S, Provenzano M, Provenza G, Di Poto A, Visai L, Arciola CR (2009) Structural and functional role of *Staphylococcus aureus* surface components recognizing adhesive matrix molecules of the host. *Future Microbiol* 4 (10):1337-1352
- Tabanella G, Nowzari H, Slots J (2009) Clinical and microbiological determinants of failing dental implants. *Clin Implant Dent Relat Res* 11 (1):24-36. doi:10.1111/j.1708-8208.2008.00088.x
- Talarico GM, Neiders ME, Comeau RL, Cohen RE (1997) Phenotypic characterization of mononuclear cells from gingiva associated with periodontitis and peri-implantitis. *J Oral Implantol* 23 (1-2):5-11
- Teronen O, Konttinen YT, Lindqvist C, Salo T, Ingman T, Lauhio A, Ding Y, Santavirta S, Sorsa T (1997) Human neutrophil collagenase MMP-8 in peri-implant sulcus fluid and its inhibition by clodronate. *J Dent Res* 76 (9):1529-1537
- Teughels W, Van Assche N, Sliepen I, Quirynen M (2006) Effect of material characteristics and/or surface topography on biofilm development. *Clin Oral Implants Res* 17 Suppl 2:68-81
- Venza I, Visalli M, Cucinotta M, De Grazia G, Teti D, Venza M (2010) Proinflammatory gene expression at chronic periodontitis and peri-implantitis sites in patients with or without type 2 diabetes. *J Periodontol* 81 (1):99-108. doi:10.1902/jop.2009.090358

- Vieira AE, Moura CC, de Souza MA, Zanetta-Barbosa D, Dechichi P (2013) Would nitric oxide be an effective marker for earlier stages of peri-implant disease? An analysis in human peri-implant sulcular fluid. *J Oral Implantol* 39 (1):37-43. doi:10.1563/AAID-JOI-D-11-00158
- Wennstrom JL, Ekestubbe A, Grondahl K, Karlsson S, Lindhe J (2004) Oral rehabilitation with implant-supported fixed partial dentures in periodontitis-susceptible subjects. A 5-year prospective study. *J Clin Periodontol* 31 (9):713-724
- Wohlfahrt JC, Aass AM, Granfeldt F, Lyngstadaas SP, Reseland JE (2014) Sulcus fluid bone marker levels and the outcome of surgical treatment of peri-implantitis. *J Clin Periodontol* 41 (4):424-431
- Xu L, Yu Z, Lee HM, Wolff MS, Golub LM, Sorsa T, Kuula H (2008) Characteristics of collagenase-2 from gingival crevicular fluid and peri-implant sulcular fluid in periodontitis and peri-implantitis patients: pilot study. *Acta Odontol Scand* 66 (4):219-224
- Zitzmann NU, Abrahamsson I, Berglundh T, Lindhe J (2002) Soft tissue reactions to plaque formation at implant abutments with different surface topography. An experimental study in dogs. *J Clin Periodontol* 29 (5):456-461
- Zitzmann NU, Berglundh T (2008) Definition and prevalence of peri-implant diseases. *J Clin Periodontol* 35 (8 Suppl):286-291
- Zitzmann NU, Berglundh T, Ericsson I, Lindhe J (2004) Spontaneous progression of experimentally induced periimplantitis. *J Clin Periodontol* 31 (10):845-849

Figure legends




Figure 1: Schematic representation of periodontal (A) and peri-implant tissues (B). The lack of periodontal ligament from the peri-implant tissues is evident. Accumulation of biofilm (marked green) on the implant surface will eventually result in inflammation of the peri-implant mucosa and establishment of peri-implant mucositis (C). Progression of this inflammation can lead to the destruction of the peri-implant tissues, including the supporting bone, culminating to peri-implantitis (D).




Figure 2: Inverted light microscopy image of a submucosal biofilm sample obtained from a site with peri-implantitis (A). Epifluorescence microscopy image of the same field, combined with fluorescence *in situ* hybridization (FISH) using a 16S rRNA-targeted oligonucleotide probe for oral *Spirochaetes* (B), oral *Synergistetes* (C), or overlap of both (D).




Figure 3: Intraoral images of a patient with peri-implantitis. The implant threads are exposed to the oral cavity due to the destruction of the supporting bone. Large biofilm deposits have been accumulated over time on the implant surface (A). Accordingly, the radiographic image of the same site demonstrates severe bone loss around the implants. In fact, left implant has no supporting bone at all (B). This implant was explanted (C), while the remaining two implants were maintained and underwent treatment that involved removal of the granulation tissue, implantoplasty and surface decontamination (D). The remission of inflammation and healing is evident already two weeks following the completion of the treatment (E). Gingival healing caps were placed on the implants, and later on the prosthetic components were adapted, so that the implant function and aesthetics are restored.